

Important:

For your safety, the nasal cannula provided for your sleep study is in an unopened package. To attach the nasal cannula:

- Open the nasal cannula package included with your kit
- Screw the plastic disc connector to the metal airflow port located on the top of the device (As pictured below)
- After 3 turns the connector will be firmly connected

